

AUG 17 1999



K991873

"This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92."

**I. Summary of Safety and Effectiveness for the
KMI™ Wrist Fusion System**

1. Submitter's name, address, telephone and fax number:

Kinetikos Medical Inc.
4115 Sorrento Valley Road
San Diego, CA 92121
Telephone: 619-558-2233 FAX: 619-558-0838

Contact Person: Michael Collins, Director of Engineering

Date Summary Prepared: May 25, 1999

2. Name of Device:

Proprietary Name: KMI Wrist Fusion System

Common/Usual Name: Plate, fixation, bone

Classification Name: Plate, fixation, bone

3. Predicate Device:

KMI Wrist Fusion System-K990094, and
Howmedica ICS Mini and Small Fragment Set-K800806

4. Description of Device:

The KMI Wrist Fusion System consists of metal plates with holes to accommodate specific screws and the screws, all manufactured of 316L stainless steel.

- 5. Intended Use:** The current KMI Wrist Fusion System (K990094) is intended for wrist arthrodesis, providing fixation of small bones such as the radius and carpal bones and indicated for use in patients suffering pain and/or loss of function due to osteoarthritis, post-traumatic arthritis, fractures, revision of failed wrist fusions, segmental bone loss, or rheumatoid arthritis, while the KMI limited wrist fusion plate is intended for use in patients suffering pain and/or loss of function due to osteoarthritis, post-traumatic arthritis, fractures, revision of failed partial wrist fusions, carpal instability, or rheumatoid arthritis. The Howmedica ICS Mini and Small Fragment Set (K800806) is used for stabilizing fractures of the small bones of the hand, foot, ankle and elbow until bony union can occur. These intended uses are substantially equivalent and do not affect safety or effectiveness.

- 6. Technological Characteristics:** The current KMI Wrist Fusion System, the KMI limited wrist fusion plate, and the Howmedica ICS Mini and Small Fragment Set consist of metal plates with holes to accommodate specific screws and the screws, all manufactured of 316L stainless steel. There are no technological characteristics that raise new issues of safety or effectiveness.

Summary of Safety and Effectiveness for the KMI™ Wrist Fusion System page 2 of 2

- 7. Summary of Performance Date: Not applicable.**
- 8. Conclusions Drawn from Nonclinical and Clinical Tests: Not applicable.**



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

AUG 17 1999

Mr. Michael Collins
Director of Engineering
Kinetikos Medical, Inc.
4115 Sorrento Valley Road
San Diego, California 92121

Re: K991873
Trade Name: KMI Wrist Fusion System
Regulatory Class: II
Product Code: HRS
Dated: June 1, 1999
Received: June 2, 1999

Dear Mr. Collins:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895.

A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

Page 2 – Mr. Micheal Collins

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or at (301) 443-6597, or at its Internet address "<http://www.fda.gov/cdrh/dsmamain.html>".

Sincerely yours,



for Celia M. Witten, Ph.D., M.D.
Director
Division of General and
Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) KMI Wrist Fusion System
KMI Medical Inc., San Diego, CA
December 1998

No 510(k) Number has been issued. K991873

Device Name: KMI Wrist Fusion System


Indications for Use:

The KMI Wrist Fusion System is comprised of the total wrist fusion plate which is intended for wrist arthrodesis, providing fixation of small bones such as the radius and carpal bones and indicated for use in patients suffering pain and/or loss of function due to osteoarthritis, post-traumatic arthritis, fractures, revision of failed wrist fusions, segmental bone loss, or rheumatoid arthritis, and the KMI limited wrist fusion plate is intended for use in patients suffering pain and/or loss of function due to osteoarthritis, post-traumatic arthritis, fractures, revision of failed partial wrist fusions, carpal instability, or rheumatoid arthritis.

(Please do not write below this line. Continue on another page if needed.)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use X OR Over-the Counter Use _____
(Per 21 CFR 801.109) (Optional format 1-2-96)


(Division Sign-Off)
Division of General Restorative Devices
510(k) Number K991873